

# Human Autotaxin immunoassay kit

Catalogue number: 31770

For the quantitative determination of autotaxin  
in serum, plasma and fluid

This package insert must be read in its entirety before using this product

Website: [www.torontobioscience.com](http://www.torontobioscience.com)

E-mail: [sales@torontobioscience.com](mailto:sales@torontobioscience.com)

## TABLE OF CONTENTS

Contents	Page
INTRODUCTION	1
PRINCIPLE OF THE ASSAY	1
REAGENTS SUPPLIED	2
OTHER MATERIALS REQUIRED, BUT NOT PROVIDED	2

STORAGE	2
PREPARATION OF REAGENTS	3
PREPARATION OF STANDARDS AND SAMPLES	4
SAMPLE HANDLING	4
ASSAY PROCEDURE	5
CALCULATION	5
TYPICAL STANDARD CURVE	6
ASSAY CHARACTERISTICS	7
REFERENCE	7
SUMMARY OF ASSAY PROCEDURE	9

## INTRODUCTION

Autotaxin, also known as ENPP-2, is a secreted glycoprotein which belongs to the ectonucleotide pyrophosphatase/phosphodiesterase (NPP) family [1,2]. Generally, NPPs can hydrolyze phosphates from nucleotides. Autotaxin exhibits the unique lysophospholipase D activity [3]. The mature protein includes two somatomedin-B-like (SMB) cysteine knot domains, a catalytic domain, and an inactive C-terminal nuclease-like domain with an EF-hand-like motif that is important in cell motility, and a region involved in autotaxin secretion [1,4,5]. There are three isoforms identified in mouse and human [6,7]. Most circulating autotaxin is the  $\beta$  form which contains 863 amino acid. Autotaxin contributes to the predominant extracellular source of the phospholipid LPA (lysophosphatidic acid) from LPC (lysophosphatidylcholine) [8-11]. Autotaxin can also produce minor amounts of sphingosine 1-phosphate and cyclic phosphatidic acid which can antagonize many of the tumorigenic properties of LPA [9,12].

Autotaxin stimulates tumor cell motility and enhances invasion and metastasis. It's upregulated in melanoma, glioblastoma, breast and lung carcinoma, follicular lymphoma and other cancers [2,8,11]. Autotaxin production by adipocytes enhances pre-adipocyte proliferation and may be elevated in obesity [11, 13].

Autotaxin is present in blood, urine, saliva, seminal and cerebrospinal fluids [2, 3]. In addition, plasma autotaxin is cleared by the liver which is elevated in liver disease [3, 14]. Normal serum or plasma autotaxin concentration is reported to be slightly higher in females than in males, and highest in pregnant females [2, 14].

## PRINCIPLE OF THE ASSAY

This assay is a quantitative sandwich ELISA. The micro-plate is pre-coated with a rabbit polyclonal antibody against human autotaxin. Standards and samples are

pipetted into the wells and any human autotaxin present is bound by the immobilized antibody. After washing away any unbound substances, a biotin labelled polyclonal antibody specific for human autotaxin is added to the wells. After wash step to remove any unbound reagents, streptavidin-HRP conjugate (STP-HRP) is added. After the last wash step, an HRP substrate solution is added and colour develops in proportion to the amount of human autotaxin bound initially. The assay is stopped and the optical density of the wells determined using a microplate reader. Since the increases in absorbance are directly proportional to the amount of captured human autotaxin, the unknown sample concentration can be interpolated from a reference curve included in each assay

### **INTENDED USE**

This human autotaxin ELISA kit is designed for quantification of human autotaxin in serum, plasma and cell culture supernatant samples.

### **REAGENTS SUPPLIED**

*Each kit is sufficient for one 96-well plate and contains the following components:*

1. Micro-titre Strips (96 wells)-Coated with a rabbit polyclonal antibody against human autotaxin, sealed.
2. 10×Wash buffer-50 ml.
3. 5×Assay buffer-30 ml.
4. 100×Detection antibody solution-A biotin labelled polyclonal antibody against human autotaxin, 0.12 ml.
5. Human autotaxin standard-50 ng of recombinant human autotaxin in a buffered protein base, lyophilised.
6. 200×STP-HRP solution- 0.06 ml.
7. Substrate solution- 12 ml, ready for use.
8. Stop solution- 12 ml, ready for use.

### **OTHER MATERIALS REQUIRED, BUT NOT PROVIDED**

1. Pipettes and pipette tips.
2. 96-well plate or manual strip washer.
3. Buffer and reagent reservoirs.
4. Paper towels or absorbent paper.
5. Plate reader capable of reading absorbency at 450 nm.
6. Distilled water or deionized water.

## **STORAGE**

The kit should be stored at 2-8°C upon receipt. Remove any unused antibody-coated strips from the micro-plate, return them to the foil pouch and re-seal. Once opened, the strips may be stored at 2-8°C for up to one month.

## **PREPARATION OF REAGENTS**

*Bring all reagents and materials to room temperature before assay.*

### **A. 1×Wash buffer.**

Prepare 1×Wash buffer by mixing the 10×Wash buffer (50 ml) with 450 ml of distilled water or deionized water. If precipitates are observed in the 10×Wash buffer bottle, warm the bottle in a 37°C water bath until the precipitates disappear. The 1×Wash buffer may be stored at 2-8°C for up to one month.

### **B. 1×Assay buffer.**

Prepare 1×Assay buffer by mixing the 5×Assay buffer (30 ml) with 120 ml of distilled water or deionized water. If precipitates are observed in the 5×Assay buffer bottle, warm the bottle in a 37°C water bath until the precipitates disappear. 1×Assay buffer may be stored for up to one month at 2-8°C.

### **C. 1×Detection antibody solution.**

Spin down the 100×Detection antibody solution briefly and dilute the desired amount of the antibody 1:100 with 1×Assay buffer, 100 µl of the 1×Detection antibody solution is required per well. Prepare only as much 1×Detection antibody solution as needed. Return the 100×Detection antibody solution to 2-8°C immediately after the necessary volume is removed.

### **D. 1×STP-HRP solution.**

Spin down the 200×STP-HRP solution briefly and dilute the desired amount of the 200×STP-HRP solution 1:200 with 1×Assay buffer, 100 µl of the 1×STP-HRP solution is required per well. Prepare only as much 1×STP-HRP solution as needed. Return the 200×STP-HRP solution to 2-8°C immediately after the necessary volume is removed.

## PREPARATION OF STANDARDS AND SAMPLES

**Human autotaxin standards:** Reconstitute the lyophilised standard with 1 ml of 1×Assay buffer to generate a standard stock solution of 50 ng/ml. Allow the standard to sit for 10 minutes with gentle agitation prior to making dilutions. Prepare serially diluted standards using 1×Assay buffer as follows:

Standard volume	Volume of 1x Assay buffer	Final concentration
50 ng/ml stock	-	50 ng/ml
250 ul of 25 ng/ml	250 ul	25 ng/ml
250 ul of 12.5 ng/ml	250 ul	12.5 ng/ml
250 ul of 6.25 ng/ml	250 ul	6.25 ng/ml
250 ul of 3.12 ng/ml	250 ul	3.12 ng/ml
250 ul of 1.56 ng/ml	250 ul	1.56 ng/ml

1×Assay buffer serves as the zero standard (0 ng/ml).

The reconstituted standard stock should be aliquoted and stored at -20°C for one month. Avoid repeating freezing/thawing cycles. Please do not store the diluted standard solutions.

## SAMPLE HANDLING

Different samples may require dilution. Serum samples requires a suggested 100-fold dilution. The certain dilution factor needs to be considered according to your requirement.

## Assay procedure

*It is recommended that all standards and samples should be run in duplicate.*

1. Add 100  $\mu\text{l}$  of standard or sample per well, incubate at room temperature for 1 hour.
2. Discard the content and tap the plate on a clean paper towel to remove residual solution in each well. Add 300  $\mu\text{l}$  of 1 $\times$ Wash buffer to each well and incubate for 1 minute. Discard the 1 $\times$ Wash buffer and tap the plate on a clean paper towel to remove residual wash buffer. Repeat the wash step for a total 3 washes.
3. Add 100  $\mu\text{l}$  of 1 $\times$ Detection antibody solution to each well, incubate at room temperature for 1 hour.
4. Wash each well 3 times as in step 2.
5. Add 100  $\mu\text{l}$  of 1 $\times$ STP-HRP solution to each well, incubate at room temperature for 20 minutes.
6. Wash each well 4 times as described in step 2.
7. Add 100  $\mu\text{l}$  of Substrate solution to each well, incubate at room temperature for 15 minutes. Protect from light.
8. Add 100  $\mu\text{l}$  of Stop solution to each well, gently tap the plate frame for a few seconds to ensure thorough mixing.
9. Measure absorbance of each well at 450 nm immediately.

### **Calculation**

1. Subtract the absorbance of the blank from that of standards and samples.
2. Generate a standard curve by plotting the absorbance obtained (y-axis) against mouse Enpp2 (ATX) concentrations (x-axis). The best fit line can be generated with any curve-fitting software by regression analysis. Any curve of 4-parameter or log-log curve fitting can be used for calculation.
3. Determine mouse Enpp2 (ATX) concentration of samples from standard curve and multiply the value by the dilution factor

### **Typical standard curve**

The following standard curve is provided for demonstration only. A standard curve should be generated for each assay.

Human autotaxin (ng/ml)	Absorbance (450 nm)	Blanked Absorbance
0	0.061	0
1.56	0.141	0.08
3.12	0.222	0.161
6.25	0.368	0.307
12.5	0.656	0.595
25	1.191	1.13
50	2.355	2.294

Human autotaxin standard curve (4-parameters)

Concentration (ng/ml)

## ASSAY CHARACTERISTICS

### A. Sensitivity

The lowest human autotaxin level that can be measured by this assay is 0.78 ng/ml.

### B. Precision

Intra-assay Precision (Precision within an assay) C.V <10%.

Inter-assay Precision (Precision between assays) C.V <10%.

### C. Spiking

Serum samples were assayed by adding 90  $\mu$ l of sample and 10  $\mu$ l of spike stock solution calculated to yield the intended 0, 5, 10 ng/ml spike concentration. The recovery of human autotaxin spiked to different levels falls in 90-110%.

### D. Linearity

To assess the linearity of the assay, samples containing and/or spiked with high concentrations of human autotaxin were serially diluted with the 1 $\times$ Assay buffer to produce samples with values within the dynamic range of the assay.

Dilution	Measured ng/ml	Expected ng/ml	Recovery %
1:40	1284	1284	100
	856.8	856.8	100
	1151	1151	100
1:80	1337.1	1284	104.1
	881	856.8	102.8
	1216	1151	105.6

## REFERENCE

1. Cimpean, Anisoara, et al. "Substrate-specifying determinants of the nucleotide pyrophosphatases/phosphodiesterases NPP1 and NPP2." *Biochemical Journal* 381.1 (2004): 71-77.
2. Okudaira, Shinichi, Hiroshi Yukiura, and Junken Aoki. "Biological roles of lysophosphatidic acid signaling through its production by autotaxin." *Biochimie* 92.6 (2010): 698-706.
3. Umezu-Goto, Makiko, et al. "Autotaxin has lysophospholipase D activity leading to tumor cell growth and motility by lysophosphatidic acid production." *J cell biol* 158.2 (2002): 227-233.
4. Hausmann, Jens, et al. "Structural basis of substrate discrimination and integrin binding by autotaxin." *Nature Structural and Molecular Biology* 18.2 (2011): 198.
5. Dennis, Jameel, et al. "Phosphodiesterase-1 $\alpha$ /autotaxin's MORFO domain regulates oligodendroglial process network formation and focal adhesion organization." *Molecular and Cellular Neuroscience* 37.2 (2008): 412-424.
6. van Meeteren, Laurens A., and Wouter H. Moolenaar. "Regulation and biological activities of the autotaxin-LPA axis." *Progress in lipid research* 46.2 (2007): 145-160.
7. Giganti, Adeline, et al. "Murine and Human Autotaxin  $\alpha$ ,  $\beta$ , and  $\gamma$  Isoforms GENE ORGANIZATION, TISSUE DISTRIBUTION, AND BIOCHEMICAL CHARACTERIZATION." *Journal of Biological Chemistry* 283.12 (2008): 7776-7789.
8. Gijbsbers, Rik, et al. "The hydrolysis of lysophospholipids and nucleotides by autotaxin (NPP2) involves a single catalytic site." *FEBS letters* 538.1-3 (2003):

60-64.

9. Tsuda, Satomi, et al. "Cyclic phosphatidic acid is produced by autotaxin in blood." *Journal of Biological Chemistry* 281.36 (2006): 26081-26088.
10. Ferry, Gilles, et al. "Functional invalidation of the autotaxin gene by a single amino acid mutation in mouse is lethal." *FEBS letters* 581.18 (2007): 3572-3578.
11. Van Meeteren, Laurens A., et al. "Autotaxin, a secreted lysophospholipase D, is essential for blood vessel formation during development." *Molecular and cellular biology* 26.13 (2006): 5015-5022.
12. Tania, Mousumi, et al. "Autotaxin: a protein with two faces." *Biochemical and biophysical research communications* 401.4 (2010): 493-497.
13. Ferry, Gilles, et al. "Autotaxin is released from adipocytes, catalyzes lysophosphatidic acid synthesis, and activates preadipocyte proliferation up-regulated expression with adipocyte differentiation and obesity." *Journal of Biological Chemistry* 278.20 (2003): 18162-18169.
14. Nakamura, Kazuhiro, et al. "Validation of an autotaxin enzyme immunoassay in human serum samples and its application to hypoalbuminemia differentiation." *Clinica chimica acta* 388.1-2 (2008): 51-58.

### **summary OF Assay procedure**

Add 100  $\mu$ l of standard or sample to each well.

Incubate at room temperature for 1 hour.

Aspirate and wash each well three times.

Add 100  $\mu$ l of 1x Detection antibody solution to each well.

Incubate at room temperature for 1 hour.

Aspirate and wash each well three times.

Add 100  $\mu$ l of 1x STP-HRP solution to each well.

Incubate at room temperature for 20 minutes.

Aspirate and wash each well four times.

Add 100  $\mu$ l of Substrate solution to each well.

Incubate at room temperature for 15 minutes.

Add 100  $\mu$ l of Stop solution to each well.

Measure absorbance of each well at 450 nm.

Calculation